

Maker	Therm.#	Main Th. Range	V <sub>0</sub>	K	Q	
R4W	8332	-2° F to 30° C	103.2	$\frac{1}{6100}$		
Temp.	Corrections		Aux.	Remarks		
	Main					
0	-.055	-.050	.0	Ica pt. Nov '62 main -.002, aux .0		
5	-.073	-.068		" " 23 Oct 64 " -.060, " .0		
10	-.078	-.073	.0	Dec 64 " -.045, " .0		
15	-.084	-.079		" Mar '66 " -.05 " .0 *		
20	-.093	-.088	-.05	Dec '66 -.05 .0		
25	-.081	-.076		Nov 67 -.05 .0		
30	-.085	-.080	-.1	DEC 68 -.05 .0		
40			-.1	MAR 71 DEC 69 -.055 .0		

Certificate #  
OBSERVER PLEASE COMMENT ON REVERSE SIDE

Maker	Therm.#	Main Th. Range	V <sub>0</sub>	K	Q	
S.V.	U 4820	-24° F to 60.4 °C	203.0	$\frac{1}{6100}$	.8672 @ 100 Kg/cm <sup>2</sup>	
		-10.5 50.5 .5			.0671 @ 400 "	
Temp.	Corrections		Aux.	Remarks		
	Main					
0	-.040	-.015	-.2	Ica pt corr. 10 MAY 63 +.025 .0		
5	-.055	-.030		DEC 63 -.010 -.1		
10	-.090	-.065	-.2	DEC 64 -.015 -.1		
15	-.090	-.065		* JAN 66 -.015 -.2		
20	-.085	-.060	-.25	DEC 66 -.015 -.2		
25	-.070	-.045		Nov '67 -.040 -.2		
30	-.105	-.080	-.2	Jan 69 -.040 -.2		
35	-.065	-.040		Dec 69 -.040 -.2		
40	-.070	-.045	-.05	DEC 70 -.040 -.2		
45	-.045	-.020				
50	-.040	-.015				
60	-.055	-.030				

Certificate #  
OBSERVER PLEASE COMMENT ON REVERSE SIDE

Appendix C. Two cards from Whitney's file recording histories of reversing thermometers.